



**Supplementary Fig. 9.** Forest plot showing that the rate of clinically acceptable accuracy of cervical screw placement on subaxial cervical segments in robot-assisted methods was 96.4% (95% CI, 84.5%–100.0%;  $I^2 = 83.735\%$ ).